

GEM Tender No:GEM/2022/B/2135779 FOR ONLINE SULPHUR ANALYZER

Replies to Pre-Bid Queries

S/N	Document No.	Page	Clause	Queries	CPCL Replies
TECHNICAL QUERIES					
1	GEM2022B2135779 CORRIGENDUM	1 of 7	1. Traces of moisture will be present in Instrument air.	We request to advise dew point of Instrument Air for correct understanding.	Normally dew point shall be in the range of -30 to -40 degC. In worst case scenario, it may go to +20 degC. Hence, bidder to design the system for worst case scenario also.
2	GEM2022B2135779 CORRIGENDUM	1 of 7	3. Cooling water shall be provided. Chloride content of cooling water is 1000 ppm, Temperature is 40 degc, pressure is (2.5 to 5.5)Kg/cm2. Refer corrigendum	We request to recheck the chloride content in cooling water. 1000 ppm chloride is very extreme and we have been facing no bid from many vendors for such application. Therefore request to kindly check chloride content again.	Chloride content provided in the corrigendum is correct and there is no change.
3	GEM2022B2135779 CORRIGENDUM	1 of 7	Traces of moisture content is present in the DHDT, ISOM and MS dual stream analyser samples. Notable content of moisture is present in DHDS analyser sample. Exact moisture content shall be provided during detailed engineering stage. Hence, bidder need to consider moisture seperators or necessary components for seperation of moisture in DHDS analyser	For DHDS Gasoil sample we request to provide estimated moisture content. Since it is notable it will change filter consideration drastically and hence request to please provide an estimate figure.	Moisture content shall be approximately in the range of 1 to 2 %.

S/N	Document No.	Page	Clause	Queries	CPCL Replies
4	GEM2022B2135779 CORRIGENDUM	3 of 7	Bidder to design the sample handling system as per operating condition. It shall withstand design conditions without damaging the sample conditioning system and analyser.	We understand that design conditions are basically upset conditions. Please note that we shall consider sample block in case condition exceeds the operating to protect from damage. We hope this is acceptable way to keep system protected from any upset above operating.	Noted. Bidder to consider suitable bypass arrangement for Normalising the upset condition.
COMMERCIAL QUERIES					
1	Additional Terms & Conditions	4 of 7	<p>9. Payment Terms 85% of the supply value within 10 days of receipt of materials at CPCL site subject to acceptance. 15% of the supply value within 10 days of completion of Site Work activities (Installation, Commissioning, SAT etc.,) and handing over of the complete system at CPCL site, duly certified by CPCL Engineer in-charge. In case site is not ready for site works, 10% of the supply value shall be released within 90 days from receipt of materials at site against furnishing an equivalent bank guarantee valid for one year period. This is apart from the submission of 3% PBG.</p>	<p>We would like to highlight that as per terms " In case site is not ready for site works, 10% of the supply value shall be released within 90 days from receipt of materials at site against furnishing an equivalent bank guarantee valid for one year period." only 10% is released against the PBG however balance amount is 15%. Therefore no clarity on how remaining 5% will be released in case of site delay. Please confirm to accept "In case site is not ready for site works, <u>15%</u> of the supply value shall be released within 90 days from receipt of materials at site against furnishing an equivalent bank guarantee valid for one year period."</p>	<p>In case site is not ready for site works, 15% of the supply value shall be released after 90 days from the date of receipt of materials at site against furnishing an equivalent bank guarantee valid for one year period. This is apart from the submission of 3% PBG. Please refer Corrigendum-II</p>

GEM Tender No:GEM/2022/B/2135779- “Procurement, Installation, Commissioning and Post Warranty Comprehensive Annual Maintenance Contract (PWCAMC) of Sulphur Analyzers”– Corrigendum - II

The following scope/specification, terms and conditions stands modified to the extent indicated under the column “Additions/Deletions/ Modifications”. All other scope/specifications, terms and conditions of the bidding document shall remain unaltered.

S/N	Tender clause No.	Page No.	Subject	Additions/Deletions/ Modifications	Bidder’s Compliance
1.	Additional Terms and Conditions (ATC)	4 of 7	9. Payment Terms	Mod: "In case site is not ready for site works, 15% of the supply value shall be released after 90 days from the date of receipt of materials at site against furnishing an equivalent bank guarantee valid for one year period. This is apart from the submission of 3% PBG"	
2.	Pre-Qualification Criteria	3 of 3	Certificate for Authorisation	Mod.: Bidder to submit Certificate of Authorisation as per the below mentioned format.	

CERTIFICATE FOR AUTHORISATION

(To be signed by Bidder’s principal’s corporate level signatory on company letter Head)

We are pleased to authorize M/s. _____ having their registered office at _____ to submit the quote on behalf of M/s. _____ for the Tender No. _____.

We, further confirm that in case where M/S. _____ fail to render their service during Supply, Warranty, PWCAMC due to some unforeseen situation, then the service support shall be provided by M/S. _____.

Name:

Mail ID:

Mobile No.:

(SIGNATURE WITH SEAL)